



**Certificate of Conformance for Quality Assurance**  
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

**K100-0102HR**  
**TOTAL HARDNESS**  
**TEST STRIPS**

**LOT#: L-5322-0216**

**EXPIRATION DATE: 10/2018**

	Date	Initials	Reference	Result	Pass	Fail
1	02/29/2016	CBP	Distilled water blank	0 ppm	X	
2	02/29/2016	CBP	Distilled water blank	0 ppm	X	
3	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
4	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
5	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
6	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
7	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
8	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
9	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	
10	02/29/2016	CBP	100mg/L Total Hardness	~100 ppm	X	

**NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT**  
 (FOR TOTAL HARDNESS REFERENCE, PASS RESULT INDICATED BY ~100PPM)  
 (FOR DISTILLED WATER BLANK, PASS RESULT INDICATED BY 0 PPM)

<b>CONTROL DATA</b>	
Manufacturer:	HACH
Part Number:	2833449
Lot Number:	A6029
Expiration Date:	01/2018

Sincerely,

**Carrie Plummer**  
 Director of Quality Assurance and Regulatory Affairs