

## Certi-Chek™ Field Verification Program • Certificate of Conformance •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. In addition to these quality assurance procedures, Certi-Chek™ testing is performed by RPC using industry accepted standard reference methods at the dialysis specific maximum allowable level for the indicated substance under test. Certi-Chek™ testing replaces the required test strip field validation called for in most Instructions for Use of non-RPC test strips. This Certi-Chek™ Field Verification Program is performed in addition to the immediate post-production QC testing that is required to confirm proper manufacturing of the test strips. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). End users of this test strip must comply with all instructions for use and storage/handling requirements.

## K100-0102 WATER HARDNESS TEST STRIPS

	100057	EVOID ATION DATE	2026 04 20
LOT#:	108067	EXPIRATION DATE:	2026-04-30

CERTI-CHEK™ FIELD VERFIFICATION TESTING									
Control	Acceptance Rate (# pass/total tested x100)	Test Value	Number of Replicates	Pass/Fail	Initials	Date			
I (Positive)	100%	1 gpg 17 ppm	13	PASS	BL	04/12/2024			

Sincerely,

Brenda Lara
Quality Manager