



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0105
MICRO-X[®] PERACETIC ACID
TEST STRIPS

LOT#: L-102002 **EXPIRATION DATE:** 2019-11-30

E-Z CHEK FIELD VERIFICATION TESTING**				
Control	Acceptance Rate* <small>(# pass/total tested x100)</small>	Pass/Fail	Initials	Date
I	100%	PASS	K.E.B.	11/7/17
II	N/A	N/A	N/A	N/A

PASS INDICATED BY ACCEPTANCE RATE OF 100% FOR EACH CONTROL TESTED

CONTROL INFORMATION		
	Control I (Pos)	Control II (Neg) <input checked="" type="checkbox"/> N/A
Manufacturer	RPC	
Part Number	MX-30004	
Lot Number	151753	
Expiration Date	2018-08-18	

NOTES: *Sample size determined per ACC-004, RPC Sampling Plan
 ** Certi-Chek™ Field Verification Testing performed according to ACC-053, K100-0105 Micro-X[®] Peracetic Acid Test Procedure

Sincerely,

Dave Cox
 Director of Quality Assurance and Regulatory Affairs