



Certificate of Conformance for Quality Assurance
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0106
E-Z CHEK[®] SENSITIVE TOTAL CHLORINE
& CHLORAMINES TEST STRIPS

LOT#: L-Z824BS **EXPIRATION DATE:** 2020-01-31

E-Z CHEK FIELD VERIFICATION TESTING**				
Control	Acceptance Rate* <small>(# pass/total tested x100)</small>	Pass/Fail	Initials	Date
I	100%	PASS	KEB	02/13/2018
II	100%	PASS	KEB	02/13/2018

PASS INDICATED BY ACCEPTANCE RATE OF 100% FOR EACH CONTROL TESTED

CONTROL INFORMATION		
	Control I (Pos)	Control II (Neg) <input type="checkbox"/> N/A
Manufacturer	EMD MILLIPORE CORP	RPC
Part Number	1.02426.0250	De-ionized Water
Lot Number	K48314926	N/A
Expiration Date	N/A	N/A

NOTES: *Sample size determined per ACC-004, RPC Sampling Plan
 ** Certi-Chek[™] Field Verification Testing performed according to ACC-008, K100-0106 E-Z Chek[®] Sensitive Total Chlorines & Chloramines Test Procedure

Sincerely,

Kathryn E. Brown
 Quality Engineer