



Certificate of Conformance for Quality Assurance
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0118/F
E-Z CHEK® ULTRA-LOW TOTAL CHLORINE
TEST STRIPS

LOT#: L-X614-0316

EXPIRATION DATE: 02/2018

| Reference | No | Date | Initials | Results | Pass | Fail |
|--------------------------------------|----|------------|----------|----------|------|------|
| 0.1ppm Chloramine T trihydrate | 1 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 2 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 3 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 4 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 5 | 03/24/2016 | CBP | 0.1+ ppm | X | |
| | 6 | 03/24/2016 | CBP | 0.1+ ppm | X | |
| | 7 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 8 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 9 | 03/24/2016 | CBP | 0.1 ppm | X | |
| | 10 | 03/24/2016 | CBP | 0.1 ppm | X | |
| Distilled Water | 11 | 03/24/2016 | CBP | 0+ ppm | X | |
| | 12 | 03/24/2016 | CBP | 0 ppm | X | |
| | 13 | 03/24/2016 | CBP | 0 ppm | X | |
| | 14 | 03/24/2016 | CBP | 0 ppm | X | |
| | 15 | 03/24/2016 | CBP | 0+ ppm | X | |
| | 16 | 03/24/2016 | CBP | 0+ ppm | X | |

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT

Specifications: 0.1ppm Chloramine T trihydrate = 0.1⁻ to 0.2⁻ ppm

Distilled Water = 0 to 0+ ppm

(The reported specification means that the result can be slightly lighter than 0.1ppm up to slightly lighter than 0.2ppm on the color chart when using a 0.1ppm standard. Likewise, for distilled water, the result can be 0ppm to slightly darker than 0ppm.)

CONTROL DATA

MANUFACTURER: EMD Millipore Corp.
 PART NUMBER: 1.02426.0250
 LOT NUMBER: K44515326 343
 EXPIRATION DATE: 7/31/2016

Sincerely,

Carrie Plummer

Director of Quality Assurance and Regulatory Affairs