



Certificate of Conformance for Certi-Chek™ QC Verification
 • *Quality Control Chart* •

This document represents certification of conformance with stocked product quality control that may be required or indicated in the Instructions for Use for performance verification of the test strip identified below. This Certi-Chek™ verification represents a QC test that is in addition to QC testing performed at the time of manufacture. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). ***With the exception of test strip QC procedures***, end users of this test strip must comply with all instructions for use and storage/handling requirements.

**K100-0118/F/BX
 ULTRA LOW TOTAL CHLORINE
 TEST STRIPS**

LOT#: L-Z604-0216

EXPIRATION DATE: 11/2017

	Date	Initials	Results	Pass	Fail
1	02/02/2016	AV	0.1PPM	X	
2	02/02/2016	AV	0.1PPM	X	
3	02/02/2016	AV	0.1PPM	X	
4	02/02/2016	AV	0.1PPM	X	
5	02/02/2016	AV	0.1PPM	X	
6	02/02/2016	AV	0.1PPM	X	
7	02/02/2016	AV	0.1PPM	X	

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT
 (PASS RESULT INDICATED BY 0.1PPM)

CONTROL DATA	
MANUFACTURER:	CHLORAMINE T
PART NUMBER:	1.02426.0250
LOT NUMBER:	K44515326-343

Sincerely,

Carrie Plummer

Director of Quality Assurance and Regulatory Affairs