



**Certificate of Conformance for Quality Assurance**  
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

**K100-0100**  
**Micro-X<sup>®</sup> RESIDUAL**  
**TEST STRIPS**

**LOT#: L-020927**

**EXPIRATION DATE: 09/2018**

	Sample	Date	Initials	0	1	3	10	30	100
Control	1	04/05/2016	AV			X			
	2	04/05/2016	AV			X			
	3	04/05/2016	AV			X			
	4	04/05/2016	AV			X			
	5	04/05/2016	AV			X			
	6	04/05/2016	AV			X			
	7	04/05/2016	AV			X			
	8	04/05/2016	AV			X			
Blank	9	04/05/2016	AV	X					
	10	04/05/2016	AV	X					

**NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT**  
 (PASS RESULT INDICATED BY X AT:  
 3 PPM FOR CONTROL; 0 PPM FOR BLANK)

<b>CONTROL DATA</b>	
Manufacturer:	Signature Care
Part Number:	NDC 21130-871-43
Lot Number:	0298582
Expiration date:	08/2017

Sincerely,

**Carrie Plummer**

Director of Quality Assurance and Regulatory Affairs