



**Certificate of Conformance for Quality Assurance**  
 • *Quality Control Chart* •

This document represents certification of conformance with stocked product quality control that may be required or indicated in the Instructions for Use for performance verification of the test strip identified below. This Certi-Chek™ verification represents a QC test that is in addition to QC testing performed at the time of manufacture. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). **With the exception of test strip QC procedures**, end users of this test strip must comply with all instructions for use and storage/handling requirements.

**K100-0101B**  
**E-Z CHEK™ RESIDUAL CHLORINE**  
**TEST STRIPS**

**LOT#: L-088016**

**EXPIRATION DATE: 07/2018**

	Date	Initials	Result	0.5	1	2	5
1	02/18/2016	AV	Pass	X			
2	02/18/2016	AV	Pass	X			
3	02/18/2016	AV	Pass	X			
4	02/18/2016	AV	Pass	X			
5	02/18/2016	AV	Pass	X			
6	02/18/2016	AV	Pass	X			
7	02/18/2016	AV	Pass	X			

**NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT**  
 (RESULTS IN PPM; PASS RESULT AT 0.5 PPM)

<b>CONTROL DATA</b>	
Manufacturer:	Spectrum
Part Number:	S1669
Lot Number:	2EE0431
Exp. Date:	06/15/2016

Sincerely,

**Carrie Plummer**

Director of Quality Assurance and Regulatory Affairs