



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with stocked product quality control that may be required or indicated in the Instructions for Use for performance verification of the test strip identified below. This Certi-Chek™ verification represents a QC test that is in addition to QC testing performed at the time of manufacture. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0101B
E-Z CHEK™ RESIDUAL CHLORINE
TEST STRIPS

LOT#: L-088016

EXPIRATION DATE: 07/2018

Table with 8 columns: ID, Date, Initials, Result, 0.5, 1, 2, 5. Rows 1-7 show 'Pass' results for 'AV' initials on '02/18/2016' with 'X' marks in the 0.5 column.

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT
(RESULTS IN PPM; PASS RESULT AT 0.5 PPM)

CONTROL DATA
Manufacturer: Spectrum
Part Number: S1669
Lot Number: 2EE0431
Exp. Date: 06/15/2016

Sincerely,

Handwritten signature of Carrie Plummer

Carrie Plummer
Director of Quality Assurance and Regulatory Affairs