



Certificate of Conformance for Quality Assurance
• Quality Control Chart •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0104
0-14 pH
TEST STRIPS

LOT#: L-45126 **EXPIRATION DATE:** 2020-12-31

E-Z CHEK FIELD VERIFICATION TESTING**				
Control	Acceptance Rate* <small>(# pass/total tested x100)</small>	Pass/Fail	Initials	Date
I	100%	Pass	AV	01/12/2017
II	N/A	N/A	N/A	N/A

PASS INDICATED BY ACCEPTANCE RATE OF 100% FOR EACH CONTROL TESTED

CONTROL INFORMATION				
	Control I-1 (Pos)	Control I-2 (Pos)	Control I-3 (Pos)	Control II (Neg) <input checked="" type="checkbox"/> N/A
Manufacturer	Hanna instruments	EMD Millipore	Hanna instruments	
Part Number	HI7004	1.09887.0001	HI7010	
Lot Number	9936	HC565541	8072	
Expiration Date	03/2021	03/12/2017	09/2019	

NOTES: *Sample size determined per ACC-004, RPC Sampling Plan
 ** Certi-Chek™ Field Verification Testing performed according to ACC-010, K100-0104 0-14 pH Test Procedure

Sincerely,

Andrea Villalobos
 RPC Quality Assurance Technician