



Certificate of Conformance for Quality Assurance
 • *Quality Control Chart* •

This document represents certification of conformance with all manufacturing, quality assurance, and test procedures required for the production and final assembly of the test strip indicated. RPC complies with all storage and handling requirements and performs quality control testing (on a lot number basis). This product was manufactured in compliance with the requirements of the Food & Drug Administration (FDA) Quality Systems Regulation (QSR). With the exception of test strip QC procedures, end users of this test strip must comply with all instructions for use and storage/handling requirements.

K100-0106
E-Z CHEK[®] SENSITIVE TOTAL CHLORINE
& CHLORAMINES
TEST STRIPS

LOT#: L-022216X

EXPIRATION DATE: 02/2018

	Date	Initials	Results	Pass	Fail
1	03/07/2016	AV	0.1PPM	X	
2	03/07/2016	AV	0.1PPM	X	
3	03/07/2016	AV	0.1PPM	X	
4	03/07/2016	AV	0.1PPM	X	
5	03/07/2016	AV	0.1PPM	X	
6	03/07/2016	AV	0.1PPM	X	
7	03/07/2016	AV	0.1PPM	X	
8	03/07/2016	AV	0.1PPM	X	
9	03/07/2016	AV	0.1PPM	X	
10	03/07/2016	AV	0.1PPM	X	

NOTE: PLACE "X" IN BOX CORRESPONDING TO RESULT
 (PASS RESULT INDICATED BY 0.1ppm)

CONTROL DATA	
MANUFACTURER:	EMD Millipore Corp.
PART NUMBER:	1.02426.0250
LOT NUMBER:	K44515326-343
EXPIRATION DATE:	07/31/2016

Sincerely,

Carrie Plummer
 Director of Quality Assurance and Regulatory Affairs