REF K100-0112

E-Z Chek® Potency Chlorine Test Strips


Instructions for Use

Indications for Use

K100-0112 E-Z Chek® Potency Chlorine Test Strips produce a semi-quantitative indication of free chlorine expressed in parts per million (ppm), for verifying 1:100 bleach solutions. E-Z Chek® Potency Chlorine Test Strips provide the capability for detection of free chlorine as recommended by the Centers for Disease Control and Prevention (CDC), of 500-615 ppm in bleach solutions. Color gradations at 100, 200, 350, 500, 750 ppm.

Test Procedure

1. Immerse test pad in sample for 2 seconds.
2. Briskly shake off excess liquid.
3. At 60 seconds, compare test pad with closest color block. Compare color at center of pad; disregard color around the edge of the indicator pad.



Warnings and Cautions

- Do not touch the test pad at the end of the strip.
- Remove only as many test strips as are required.
- Keep all unused test strips in the original bottle with lid tightly closed. Replace cap immediately.
- Do not allow the test strips to contact work surfaces that may be contaminated with interfering substances.
- Do not leave the test strips in areas exposed to chlorine vapors or other oxidizing vapors.
- Expiration date and lot number are printed on the container label.
- Verify that the test strips are within their expiration date before use, and do not use them after the expiration date.
- Do not remove the desiccant from cap/lid
- The test strip container stopper (cap/lid) contains a non-toxic drying agent. If swallowed, drink plenty of water.

Interferences

No interferences are known.

Chemical Properties of the Test

OSHA classifies test strips as an article not requiring a Safety Data Sheet (SDS).

Quality Control

RPC performs an independent quality control test on each lot number of test strips (Certi-Chek™ Field Verification Program). RPC includes a certificate of conformance with each shipment per lot and specifies the IFU revision applicable. Certificates of conformance are also available upon request and/or can be downloaded at www.rpc-rabrenco.com.

REF **K100-0112**

E-Z Chek® Potency Chlorine Test Strips

Instructions for Use

Storage and Handling

- **Symbol Legend:**

Temperature Limit	Keep Dry	Keep Away from Sunlight	Test strip per bottle	Recyclable	Immediately close after opening
